

CERTIFICATE

Survey of 27 June 2025

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Virus Immunology - Epstein Barr Virus (Ab) (352):

Validity 12 months:

Anti-EBNA 1 IgG

Anti-EBNA IgM

Anti-VCA IgG

Anti-VCA IgM

Strip immunoblot anti-EBV IgG

Strip immunoblot anti-EBV IgM

Anti-VCA IgG - avidity

Clinical interpretation

(R) analysis is subject to the RiliBÄK

Customer:

10011570

TestLine Clinical Diagnostics s.r.o.

Krizikova 188/68

61200 Brno



Düsseldorf, 16 July 2025



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. Holger F. Rabenau
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 27 June 2025

You have participated in the External Quality Assessment with the following analysis

Virus Immunology - Epstein Barr Virus (Ab) (352):

Anti-EBNA 1 IgG
Anti-EBNA IgM
Anti-EA IgG
Anti-EA IgM
Anti-VCA IgG
Anti-VCA IgA
Anti-VCA IgM
Strip immunoblot anti-EBV IgG
Strip immunoblot anti-EBV IgA
Strip immunoblot anti-EBV IgM
Anti-VCA IgG - avidity
Clinical interpretation

(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno



Düsseldorf, 16 July 2025



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. Holger F. Rabenau
(Adviser)



Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 27 June 2025

Adviser:

Prof. Dr. Holger F. Rabenau
Institut für Medizinische Virologie
Universitätsklinikum Frankfurt
Paul-Ehrlich-Str. 40
60596 Frankfurt

Tel.: +49 211-1592 13 0

Mail: holger.rabenau@instand-ringversuch.de

352

Virus Immunology - Epstein Barr Virus (Ab)

Anti-EBNA 1 IgG

+

Name of test	Unit	Method	Manufacturer				
EIA EBV EBNA-1 IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	positive (3) negative (1)	positive (3) negative (1)		SV SV		+ +
Name of test	Unit	Method	Manufacturer				
CLIA EBV EBNA-1 IgG	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A- 250610	Your value	Lower limit	Upper limit	TV- Type	Target value	Meets criteria
	-01 -02	46.5 4.18					
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	positive (3) negative (1)	positive (3) negative (1)		SV SV		+ +

Anti-EBNA IgM

+

Name of test	Unit	Method	Manufacturer				
EIA EBV EBNA-1 IgM		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	negative (1) negative (1)	negative (1) negative (1)		SV SV		+ +
Name of test	Unit	Method	Manufacturer				
CLIA EBV EBNA-1 IgM	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A- 250610	Your value	Lower limit	Upper limit	TV- Type	Target value	Meets criteria
	-01 -02	7.30 5.65					
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	negative (1) negative (1)	negative (1) negative (1)		SV SV		+ +

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 27 June 2025

Anti-EA IgG

Name of test	Unit	Method	Manufacturer				
EIA EBV EA-D IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	negative (1)					Sample not evaluated
	-02	negative (1)					Sample not evaluated
Name of test	Unit	Method	Manufacturer				
CLIA EBV EA-D IgG	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A-250610	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	-01	27.6					Sample not evaluated
	-02	5.12					Sample not evaluated
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	positive (3)					Sample not evaluated
	-02	negative (1)					Sample not evaluated

Anti-EA IgM

Name of test	Unit	Method	Manufacturer				
EIA EBV EA-D IgM		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	negative (1)					Sample not evaluated
	-02	negative (1)					Sample not evaluated
Name of test	Unit	Method	Manufacturer				
CLIA EBV EA-D IgM	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A-250610	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	-01	4.49					Sample not evaluated
	-02	4.47					Sample not evaluated
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	negative (1)					Sample not evaluated
	-02	negative (1)					Sample not evaluated

Anti-VCA IgG

Name of test	Unit	Method	Manufacturer				
EIA EBV VCA IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	positive (3)	positive (3)		SV		+
	-02	negative (1)	negative (1)		SV		+

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 27 June 2025

Anti-VCA IgG

+

Name of test	Unit	Method	Manufacturer				
CLIA EBV VCA IgG	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A-250610	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	-01	313					
	-02	3.00					
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	positive (3)	positive (3)		SV		+
	-02	negative (1)	negative (1)		SV		+

Anti-VCA IgA

Name of test	Unit	Method	Manufacturer				
EIA EBV VCA IgA		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	negative (1) negative (1)					
Name of test	Unit	Method	Manufacturer				
CLIA EBV VCA IgA	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A- 250610	Your value	Lower limit	Upper limit	TV- Type	Target value	Meets criteria
	-01 -02	5.58 2.98					
Result (qual)	Sample 352A- 250610	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01 -02	negative (1) negative (1)					

Anti-VCA IgM

+

Name of test	Unit	Method	Manufacturer				
EIA EBV VCA IgM		Immunological test (ELISA) (190)	TEST-LINE (TT)				
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	negative (1)	negative (1)		SV		+
	-02	negative (1)	negative (1)		SV		+
Name of test	Unit	Method	Manufacturer				
CLIA EBV VCA IgM	U/ml	Immunological test (CLIA) (228)	TEST-LINE (TT)				
Result (quant)	Sample 352A-250610	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	-01	7.44					
	-02	4.09					
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	-01	negative (1)	negative (1)		SV		+
	-02	negative (1)	negative (1)		SV		+

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 27 June 2025

Strip immunoblot anti-EBV IgG

+

Name of test	Unit	Method	Manufacturer			
BLOT-LINE EBV IgG		STRIP-IA (381)	TEST-LINE (TT)			
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	-01	positive (3)	positive (3)	SV		+
	-02	negative (1)	negative (1)	SV		+

Strip immunoblot anti-EBV IgA

Name of test	Unit	Method	Manufacturer			
BLOT-LINE EBV IgA		STRIP-IA (381)	TEST-LINE (TT)			
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	-01	negative (1)				
	-02	negative (1)				

Sample not evaluated
Sample not evaluated

Strip immunoblot anti-EBV IgM

+

Name of test	Unit	Method	Manufacturer			
BLOT-LINE EBV IgM		STRIP-IA (381)	TEST-LINE (TT)			
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	-01	negative (1)	negative (1)	SV		+
	-02	negative (1)	negative (1)	SV		+

Anti-VCA IgG - avidity

+

Name of test	Unit	Method	Manufacturer			
EBV VCA IgG Avidität		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	-01	high (1)	high (1)	SV		+
	-02	no avidity (4)	no avidity (4), not done (98)	SV		+
Avidity in percent	Sample 352A-250610	Your value				
	-01	74.0				
	-02					

Clinical interpretation

+

Name of test	Unit	Method	Manufacturer			
Result (qual)	Sample 352A-250610	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	-01	evidence for a past infection (2)	evidence for a past infection (2)	SV		+
	-02	no evidence for infection (1)	no evidence for infection (1)	SV		+

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".