

CERTIFICATE

Survey of 22 November 2024

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Virus Immunology - Parvovirus B19 (Ab) (342):

Validity 12 months:


Anti-parvovirus B19 IgG

Anti-parvovirus B19 IgM


(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 19 December 2024



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Martin Enders
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 22 November 2024

You have participated in the External Quality Assessment with the following analysis

Virus Immunology - Parvovirus B19 (Ab) (342):


Anti-parvovirus B19 IgG

Anti-parvovirus B19 IgM


(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 19 December 2024



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Martin Enders
(Adviser)

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 22 November 2024

Adviser:

Prof. Dr. med. Martin Enders
Labor Prof. G. Enders & Kollegen, MVZ
Medizinische Diagnostik
Rosenbergstr. 85
70193 Stuttgart

Tel.: +49 711-6357 120
Mail: martin.enders@instand-ringversuch.de

342

Virus Immunology - Parvovirus B19 (Ab)

Anti-parvovirus B19 IgG

Name of test	Unit	Method	Manufacturer			
EIA Parvovirus B19 IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	342A-241105-01	negative (1)	negative (1)	SV		+
	342A-241105-02	positive (3)	positive (3)	SV		+
	342A-241105-03	positive (3)	positive (3)	SV		+
	342A-241105-04	positive (3)	positive (3)	SV		+
Name of test	Unit	Method	Manufacturer			
CLIA Parvovirus B19 IgG		Immunological test (CLIA) (228)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	342A-241105-01	negative (1)	negative (1)	SV		+
	342A-241105-02	positive (3)	positive (3)	SV		+
	342A-241105-03	positive (3)	positive (3)	SV		+
	342A-241105-04	positive (3)	positive (3)	SV		+

Anti-parvovirus B19 IgM

Name of test	Unit	Method	Manufacturer			
EIA Parvovirus B19 IgM		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	342A-241105-01	negative (1)	negative (1)	SV		+
	342A-241105-02	negative (1)	negative (1)	SV		+
	342A-241105-03	negative (1)	negative (1)	SV		+
	342A-241105-04	negative (1)	negative (1)	SV		+
Name of test	Unit	Method	Manufacturer			
CLIA Parvovirus B19 IgM		Immunological test (CLIA) (228)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	342A-241105-01	negative (1)	negative (1)	SV		+
	342A-241105-02	negative (1)	negative (1)	SV		+
	342A-241105-03	negative (1)	negative (1)	SV		+
	342A-241105-04	negative (1)	negative (1)	SV		+

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".