

CERTIFICATE

Survey of 22 November 2024

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Virus Immunology - Rubella Virus (Ab) (341):

Validity 12 months:

Anti-rubella virus IgG (R: B2)

Anti-rubella virus IgG - avidity (R: B2)

Anti-rubella virus IgM (R: B2)


(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 19 December 2024



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Martin Enders
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 22 November 2024

You have participated in the External Quality Assessment with the following analysis

Virus Immunology - Rubella Virus (Ab) (341):

Anti-rubella virus IgG (R: B2)

Anti-rubella virus IgG - avidity (R: B2)

Anti-rubella virus IgM (R: B2)


(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 19 December 2024



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Martin Enders
(Adviser)

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 22 November 2024

Adviser:

Prof. Dr. med. Martin Enders
Labor Prof. G. Enders & Kollegen, MVZ
Medizinische Diagnostik
Rosenbergstr. 85
70193 Stuttgart

Tel.: +49 711-6357 120
Mail: martin.enders@instand-ringversuch.de

341

Virus Immunology - Rubella Virus (Ab)

Anti-rubella virus IgG

+

Name of test	Unit	Method	Manufacturer			
EIA Rubella IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	341A-241105-01	positive (3)	positive (3)	SV		+
	341A-241105-02	negative (1)	negative (1)	SV		+
Name of test	Unit	Method	Manufacturer			
CLIA Rubella IgG		Immunological test (CLIA) (228)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	341A-241105-01	positive (3)	positive (3)	SV		+
	341A-241105-02	negative (1)	negative (1)	SV		+

Anti-rubella virus IgG - avidity

+

Name of test	Unit	Method	Manufacturer			
EIA Rubella IgG		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	341A-241105-01	high (1)	high (1)	SV		+
	341A-241105-02	no avidity (4)	no avidity (4), not done (98)	SV		+
Avidity in percent	Sample	Your value				
	341A-241105-01	93.0				
	341A-241105-02					

Anti-rubella virus IgM

+

Name of test	Unit	Method	Manufacturer			
EIA Rubella IgM		Immunological test (ELISA) (190)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	341A-241105-01	negative (1)	negative (1)	SV		+
	341A-241105-02	negative (1)	negative (1)	SV		+
Name of test	Unit	Method	Manufacturer			
CLIA Rubella IgM		Immunological test (CLIA) (228)	TEST-LINE (TT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	341A-241105-01	negative (1)	negative (1)	SV		+
	341A-241105-02	negative (1)	negative (1)	SV		+

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".